



Certificate of Traceability

Manufacturer Certification

This certifies that **TaiDoc Technology Corporation**, using **Immunoassay on TD-4611** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13**

The system evaluated was:

Instrument: TD-4611	Calibrator Lot: 609103, 609103
Reagent Lot: 17I201	Calibrator Assigned Values: 5.4%, 11.7%

Date of Certification: January 1, 2018

Certification Expires: January 1, 2019

David Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Chy

SRL director/ supervisor